

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 10058 (1985): Forceps, Fixation, Graefe's Pattern
(Modified) [MHD 5: Ophthalmic Instruments and Appliances]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS : 10058 - 1985

Indian Standard
SPECIFICATION FOR
FORCEPS, FIXATION, GRAEFE'S
PATTERN (MODIFIED)
(*First Revision*)

UDC 615.472.4 : 617.7-089-72



© Copyright 1986

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Indian Standard

SPECIFICATION FOR FORCEPS, FIXATION, GRAEFE'S PATTERN (MODIFIED) (*First Revision*)

Eye Surgery Instruments Sectional Committee, CPDC 27

Chairman

DR K. LALL

Representing

Ministry of Health (DGHS), New Delhi

Members

DR P. M. KAPOOR (*Alternate to*
Dr K. Lall)

SHRI ARYA MUNI
COL M. S. BOPARAI
DR D. B. CHANDRA

Arya Muni & Co, Amritsar
Ministry of Defence (DGAEMS), New Delhi
Department of Health and Family Welfare, Govern-
ment of U.P., Lucknow
Department of Health and Family Welfare, Govern-
ment of Punjab, Chandigarh

DR DALJIT SINGH

DR CHARANJIT LAL (*Alternate*)

SHRI K. H. GANDHI

Ministry of Defence (DGI)

SHRI V. K. BAJPAI (*Alternate*)

DR A. K. GUPTA

Delhi Administration (M & PH Department), Delhi
Kalelker Surgical Industries, Bombay

SHRI P. Y. KALELKER

SHRI Y. P. KALELKER (*Alternate*)

DR A. KRISHNA MURTHY

Department of Health and Family Welfare, Govern-
ment of Karnataka, Bangalore

DR H. THIMMAIAH (*Alternate*)

SHRI A. K. MAITI

Directorate of Industries, Government of West
Bengal, Calcutta

SHRI S. GUPTA (*Alternate*)

SHRI K. MANIVANNAN

Directorate of Industries, Government of Haryana,
Chandigarh

DR I. P. SINGH PARMAR (*Alternate*)

DR D. L. MARIA

Directorate of Medical Education and Research,
Bombay

DR S. N. PATEL (*Alternate I*)

DR S. K. DHAWAN (*Alternate II*)

(*Continued on page 2*)

© Copyright 1986

INDIAN STANDARDS INSTITUTION

This publication is protected under the *Indian Copyright Act* (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

(Continued from page 1)

<i>Members</i>	<i>Representing</i>
SHRI G. S. MATHARU	Indo-German Surgical Corporation, Bombay
SHRI MOHINDER LAL	Punjab Surgical Instruments Manufacturers & Suppliers Association, Jalandhar
SHRI SUBIT MUKERJI	H. Mukerji & Banerjee Surgicals Pvt Ltd, Calcutta
SHRI SUDHIR MUKERJI (Alternate)	
DR R. MUSALAPPA REDDY	Municipal Administration Department, Government of Andhra Pradesh, Hyderabad
SHRI C. NARAYANA RAO	Bharat Electronics Ltd, Machilipatnam
SHRI A. M. JOSE (Alternate)	
SHRI SURINDER NAYYAR	Hansraj Nayyar (Surgical) Pvt Ltd, Amritsar
SHRI VIJAY NAYYAR (Alternate)	
SHRI B. L. PAHWA	Surgical Manufacturers & Traders Association, Delhi
SHRI BALWANT SINGH (Alternate)	
SHRI B. L. PAHWA	Speedway Surgical Company, Delhi
SHRI VIPIN PAHWA (Alternate)	
SHRI K. L. PAHWA	Speedway Surgical Company, Ludhiana
SHRI P. S. PAHWA (Alternate)	
DR S. SATHIAVAKESAN	Health and Family Welfare Department, Government of Tamil Nadu, Madras
PROF (DR) P. SIVA REDDY	All India Ophthalmological Society, Madras
DR J. AGARWAL (Alternate)	
DR K. S. SUBRAHMANYAM	Health and Family Welfare Department, Government of Kerala, Trivandrum
DR M. V. JOHN (Alternate)	
SHRI P. R. T. THULASINGAM	Surgical Instruments Plant, IDPL, Madras
SHRI A. K. THANIKACHALAM (Alternate)	
DR A. S. SETHI,	Director General, ISI (Ex-officio Member)
Director (Consr Prods & Med Instrs)	

Secretary

SHRI M. K. BHATIA
Deputy Director (Consr Prods & Med Instrs), ISI

Indian Standard

SPECIFICATION FOR FORCEPS, FIXATION, GRAEFE'S PATTERN (MODIFIED) (*First Revision*)

0. FOREWORD

0.1 This Indian Standard (First Revision) was adopted by the Indian Standards Institution on 25 November 1985, after the draft finalized by the Eye Surgery Instruments Sectional Committee had been approved by the Consumer Products & Medical Instruments Division Council.

0.2 This standard was first published in 1981. The present revision has been taken up to include certain modifications in order to bring the specification in line with the modern manufacturing practices. In the revised standard the non-functional dimensions have been rounded off. The tolerance for the functional dimensions have been modified keeping in view the present manufacturing technique. A clause on sampling and inspection has been added. The test methods have been brought in line with the test methods covered in other similar standards of eye surgery instruments.

0.3 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS : 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard covers requirements for Graefe's pattern fixation forceps used in eye surgery.

2. SHAPE AND DIMENSIONS

2.1 The shape and dimension shall be as shown in Fig. 1.

2.2 Joint — Shall conform to requirements given in 5 of IS : 3642-1978†.

*Rules for rounding off numerical values (*first revision*).

†General requirements for surgical instruments (*first revision*).




FIG. 1 FORCEPS, FIXATION, GRAEFE'S PATTERN

2.3 Teeth — Shall conform to requirements given in Section 4 of IS : 3642-1978* combination being 1 in 2, 2 in 3, 3 in 4, or 5 in 6.

2.4 The tolerance on dimensions of working ends shall conform to 'fine' class of deviation and the tolerance for remaining dimensions shall conform to 'medium' class of deviations of IS : 2102 (Part 1)-1980†.

3. MATERIAL

3.1 The material for forceps shall be stainless steel conforming to designation 30Cr 13 of IS : 6603-1972‡.

4. REQUIREMENTS

4.1 The forceps shall be finished smooth and shall be free from burrs, pits, cracks, and other surface defects.

4.2 The edges shall be even and rounded off.

4.3 The teeth shall be clear, clean and match perfectly.

4.4 The forceps shall be matt finished and passivated.

5. HARDNESS

5.1 The forceps shall be hardened and tempered to give a hardness of 400 to 460 HV.

6. TESTS

6.1 Test for Engagement of Teeth — A latex sheet of 0.05 mm thickness conforming to IS : 5430-1981§ be stretched over the tip of the fingers. The stretched latex sheet shall then be gripped by the teeth of the forceps and pulled to a distance of 6 mm. The forceps shall hold the latex sheet firmly without any tendency to slipping and the latex sheet shall not be punctured or damaged.

6.2 Flexibility Test — The arms of the forceps after maximum closure by manual compression shall not take a permanent set and the teeth shall continue to engage and disengage accurately and properly.

6.2.1 By the application of force at the tip of one arm, the arm of the forceps shall be deflected in a plane at right angles to the plane of the arm, by 20 mm as measured at the tip of the forceps. On release of the force no permanent set shall be observed. The test shall be repeated on the other arm.

*General requirements for surgical instruments (*first revision*).

†General tolerances for dimensions and form and position: Part 1 General tolerances for linear and angular dimensions (*second revision*).

‡Specification for stainless steel bars and flats.

§Specification for ammonia preserved concentrated natural rubber latex (*first revision*).

6.3 Corrosion Resistance Test — The forceps shall satisfy boiling and autoclaving test as specified in IS : 7531-1975*.

6.4 Load Closure Test — The load $215\text{ g} \pm 20\text{ g}$ when applied at a distance of 45 mm from the tips of the jaws shall make the teeth to just interlock.

7. SAMPLING AND CRITERIA FOR CONFORMITY

7.1 A suitable sampling scheme and criteria for acceptance is given in Appendix A.

8. MARKING

8.1 Each instrument shall be marked with the manufacturer's name, initials or recognized trade-mark and the word 'stainless steel' or letters 'SS'.

8.1.1 The forceps may also be marked with ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The ISI Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products are also continuously checked by ISI for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

9. PACKING

9.1 As agreed to between the purchaser and the supplier.

A P P E N D I X A

(Clause 7.1)

SAMPLING AND CRITERIA FOR CONFORMITY FOR FORCEPS

A-1. LOT

A-1.1 In any consignment, all the forceps produced from the same material, shape and dimensions under similar conditions, shall constitute a lot.

A-2. The number of forceps to be selected from each lot shall depend upon the size of the lot and shall be in accordance with col 1 and 2 of Table 1.

*Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments.

TABLE 1 SCALE OF SAMPLING
(Clause A-2)

LOT SIZE	SAMPLE SIZE
(1)	(2)
Up to 15	2
16 „ 50	3
51 „ 150	5
151 and above	8

A-2.1 These forceps shall be selected from the lot at random and in order to ensure the randomness of selection, procedures given in IS : 4905-1968* may be followed.

A-3. NUMBER OF TESTS AND CRITERIA FOR CONFORMITY

A-3.1 All the forceps selected at random in accordance with col 1 and 2 of Table 1 shall be tested for hardness, requirements load closure, engagement of teeth, flexibility, and corrosion resistance. The lot shall be considered as conforming to these requirements if none of the forceps in the sample is found to be defective in any of these tests.

*Methods for random sampling.

INTERNATIONAL SYSTEM OF UNITS (SI UNITS)

Base Units

QUANTITY	UNIT	SYMBOL
Length	metre	m
Mass	kilogram	kg
Time	second	s
Electric current	ampere	A
Thermodynamic temperature	kelvin	K
Luminous intensity	candela	cd
Amount of substance	mole	mol

Supplementary Units

QUANTITY	UNIT	SYMBOL
Plane angle	radian	rad
Solid angle	steradian	sr

Derived Units

QUANTITY	UNIT	SYMBOL	DEFINITION
Force	newton	N	1 N = 1 kg.m/s ²
Energy	joule	J	1 J = 1 N.m
Power	watt	W	1 W = 1 J/s
Flux	weber	Wb	1 Wb = 1 V.s
Flux density	tesla	T	1 T = 1 Wb/m ²
Frequency	hertz	Hz	1 Hz = 1 c/s (s ⁻¹)
Electric conductance	siemens	S	1 S = 1 A/V
Electromotive force	volt	V	1 V = 1 W/A
Pressure, stress	pascal	Pa	1 Pa = 1 N/m ²